

Pump Settings

Select:

Options > My Pump > Personal Profiles > Pump Settings

Quick Bolus Off On

Increment Type Units of Insulin Grams of Carbohydrates

Increment Amount 0.5u 1u 2u 5u
 2g 5g 10g 15g

Max Bolus units

Basal Limit u/hr

Control-IQ

Select:

Options > My Pump > Control-IQ

Control-IQ Off On

To turn on Control-IQ, you need to enter your weight & Total Daily Insulin. The weight entered must be no less than 24.9kg or 55 lbs and the Total Daily Insulin must be between 10 - 100 units/day.

Weight kg lbs

kg lbs

Total Daily Insulin units

Sleep Activity Schedules (Control-IQ Technology Only)

Select:

Options > Activity > Sleep Schedules

Sleep Schedules	On/Off	Selected Days	Start Time	End Time
1				
2				

Pump Reminder Settings

Select:

Options > My Pump > Alerts & Reminders > Pump Reminders

Low BG Off On

High BG Off On

Remind Me Below mmol/L

Remind Me Above mmol/L

Remind Me After mins/hrs

Remind Me After mins/hrs

After Bolus BG Off On

Remind Me After

mins/hrs

Missed Meal Bolus

Reminder	On/Off	Selected Days	Start Time	End Time
1				
2				
3				
4				

Pump Alert Settings

Select:

Options > My Pump > Alerts & Reminders > Pump Alerts**Low Insulin** units**Auto-Off** Off OnTime hours**Device Settings**

Select:

Options > Device Settings**Display Settings****Screen Time Out** 15 secs 30 secs 60 secs 120 secs**Time and Date** Time and Date set**Sound Volume****Pump Alarms** High Medium Low Vibrate**Pump Alerts** High Medium Low Vibrate**Reminders** High Medium Low Vibrate**Keypad** High Medium Low Vibrate**Bolus** High Medium Low Vibrate**Quick Bolus** High Medium Low Vibrate**Fill Tubing** High Medium Low Vibrate**CGM Alerts** Vibrate Soft Normal HypoRepeat**Security PIN** Off On PIN

My CGM

Select:

Options > My CGM > CGM Alerts

High and Low

High Alert Off On **Alert Me Above** mmol/L

Repeat Never 15 mins 30 mins 1 hr
 2 hrs 3 hrs 4 hrs 5 hrs

Low Alert Off On **Alert Me Below** mmol/L

Repeat Never 15 mins 30 mins 1 hr
 2 hrs 3 hrs 4 hrs 5 hrs

Rise and Fall

Rise Alert Off On

Rate 0.11 mmol/L/min 0.17 mmol/L/min

Fall Alert Off On

Rate 0.11 mmol/L/min 0.17 mmol/L/min

Out of Range Off On **Alert Me After** mins/hours

Select:

Options > My CGM > Transmitter ID

Transmitter ID

Site Reminder

Select:

Options > Load > Site Reminder

Site Reminder Off On

Remind Me In 1 day 2 days 3 days

Remind Me At

For more information and support on the t:slim X2 insulin pump, please contact the AMSL Diabetes Technical Support Team on **1300 851 056**.

amsldiabetes.com.au    



Always read the label and use only as directed. Systems like the t:slim X2 insulin pump with Control-IQ technology are not substitutes for the active management of diabetes. Control-IQ technology should not be used by anyone under the age of six years old. It should also not be used in patients who require less than 10 units of insulin per day or who weigh less than 25 kgs. The Control-IQ feature relies on current Dexcom G6 CGM (sold separately) readings to function and will not be able to predict glucose levels and suspend insulin delivery if a user's CGM is not functioning properly or their pump is unable to receive the CGM signal. Users should always pay attention to their symptoms, actively monitor and manage glucose levels and treat according to their healthcare professionals' recommendations. Fast acting insulin is required for the t:slim X2 insulin pump, consult a healthcare professional to see which product is right for you. The Bluetooth® word mark and logos are registered trade-marks owned by the Bluetooth SIG, Inc. and any use of such marks by Tandem Diabetes Care is under license. All other trademarks and copyrights are the property of their respective owners. AMSL is a subsidiary of Dexcom. ARTG 304681. PR-100-467 January 2023